Calibration of the Tango® Exercise Stress BP Monitor

It is recommended that the monitor be calibrated annually to verify the accuracy of the pressure transducers and indicators.

1. USE THE "VERIFY CALIBRATION" MODE
in the main menu view under "UTILITIES."

2. CONNECT A MERCURY MANOMETER
with a bulb to the pneumatic connector using a T-connector, SunTech part # 750 0018 ATA.

3. VERIFY THAT THE CUFF PRESSURE
on the display is within ±2mmHg of the pressure value on the mercury manometer through a range of pressures from 0-300mmHg. If not, contact SunTech about calibration.